



## Drug Recall

**DATE OF RECALL:** February 12, 2020

**DRUG NAME:** Minimed Insulin Pump

**RECALLING FIRM:** Medtronic

**REASON FOR RECALL:** This recall is being issued due to a missing or broken retainer ring which helps to lock the insulin cartridge into place in the pump's reservoir compartment. If the cartridge is not locked firmly into place, under or over delivery of insulin may occur, which could result in hypoglycemia or hyperglycemia.

**FDA LINK:** N/A

**OTHER DETAILS:** N/A

### RECALLED PRODUCT:

Product Description	NDC Number	Affected Lots/Expire Date
MiniMed 630G Guardian Pre MIS Transmit	43169-0800-40	All lots before October 2019
MiniMed 630G Insulin Pump Kit	43169-0752-72	All lots before October 2019
MiniMed 670G Insulin Pump MIS	43169-0946-04	All lots before August 2019
MiniMed 670G Insulin Pump MIS	76300-0278-42	All lots before August 2019