



Drug Recall

DATE OF RECALL: April 21, 2020

DRUG NAME: PosiFlush Syringe 10 mL saline flush syringes

RECALLING FIRM: BD

REASON FOR RECALL: This recall was issued because this product has been confirmed to exhibit holes in the packaging, which impacts package integrity and potentially compromises a sterile field.

FDA LINK: N/A

OTHER DETAILS: N/A

RECALLED PRODUCT:

Product Description	NDC Number	Affected Lots/Expire Date
PosiFlush Syringe 10 mL saline flush syringes	08290-3065-53	8353952 exp. 12/31/21 9011582 exp. 12/31/21 9017875 exp. 12/31/21 9024676 exp. 01/31/22 9045702 exp. 01/31/22 9060999 exp. 02/28/22 9079716 exp. 02/28/22 9127571 exp. 02/28/22 9143529 exp. 04/30/22 9156595 exp. 05/31/22 9163601 exp. 05/31/22