



Drug Recall

DATE OF RECALL: May 07, 2020

DRUG NAME: True Metrix Air Blood Glucose Meter

RECALLING FIRM: Trividia Health, Inc.

REASON FOR RECALL: This recall is being issued because the single affected meter has an incorrect factory-set unit of measure; the meter displays glucose results in mmol/L rather than mg/dL. If the consumer does not notice the incorrect unit of measure, it is possible that the meter glucose result will be read as a lower blood glucose result than expected, and this may result in the patient's glucose level remaining high, which can lead to serious injury or impairment with risk of death.

FDA LINK: <https://www.fda.gov/safety/recalls-market-withdrawals-safety-alerts/trividia-health-inc-issues-nationwide-voluntary-recall-isolated-true-metrix-air-blood-glucose-meter>

OTHER DETAILS: See FDA link.

RECALLED PRODUCT:

Product Description	Serial Number	Affected Lots
TRUE METRIX Blood Glucose Meter Kit	TA1548753	KW0135